

Alcotest 7110 Calibration Record

Equipment

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00665	Calib. Date:	04/04/2017
Certification File No.:	00630	Cert. No.:	00026
Linearity File No.:	00631	Cert. Date:	10/21/2016
Solution File No.:	00664	Lin. Date:	10/21/2016
Sequential File No.:	00665	Soln. Date:	04/03/2017
		File Date:	04/04/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0217
Solution Control Lot:	16220	Expires:	09/07/2018
		Bottle No.:	1346

Coordinator

Last Name:	KLIMIK	First Name:	DAVID	MI:	W
Signature:	<u><i>David D. Klimik</i></u> #7040			Badge No.:	7040
				Date:	04/04/2017

*Black Key Temperature Probe Serial.....# DDVJP2-003 (DK)

*Digital NIST Temperature Measuring System Serial.....# 150649679 (DK)

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part I - Control Tests

Equipment

Alcotest 7110 MKIII-C
Location: LONG BRANCH POLICE DEPT
Serial No.: ARXA-0056
Calibration File No.: 00665
Calib. Date: 04/04/2017
Calib. No.: 00026
Certification File No.: 00666
Cert. Date: 04/04/2017
Cert. No.: 00020
Linearity File No.: 00631
Lin. Date: 10/21/2016
Lin. No.: 00019
Solution File No.: 00664
Soln. Date: 04/03/2017
Soln. No.: 00136
Sequential File No.: 00666
File Date: 04/04/2017
Calibrating Unit: WET
Model No.: CU-34
Serial No.: DDXN S3-0217
Control Solution %: 0.100%
Expires: 09/07/2018
Solution Control Lot: 16220
Bottle No.: 1346

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:38D		
Control 1 EC	0.099%	12:39D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:39D		
Control 2 EC	0.099%	12:40D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:41D		
Control 3 EC	0.098%	12:42D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:43D		

All tests within acceptable tolerance.

Coordinator

Last Name: KLIMIK

First Name: DAVID

MI: W

Signature: David Klimik

Badge No.: 7040

Date: 04/04/2017

Pursuant to law, and the "Chemical Breath Testing Regulations" N.J.A.C. 13:51, I am a duly appointed Breath Test Coordinator/Instructor. In my official capacity, and consistent with "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist of the Division of State Police, I perform calibration checks on approved instruments employing infrared analysis and electrochemical analysis, when utilized in a single approved instrument as a dual system of chemical breath testing. Pursuant to, and consistent with, the current "Calibration Check Procedure for Alcotest 7110," as established by the Chief Forensic Scientist, I performed a Calibration Check on the approved instrument identified on this certificate. The results of my Calibration Check are recorded on this certificate, which consists of two parts on two pages: Part I - Control Tests; and Part II - Linearity Tests. I certify that the foregoing statements made by me are true. I am aware that if any of the foregoing statements made by me are wilfully false, I am subject to punishment.

Alcotest 7110 Calibration Certificate

Part II - Linearity Tests

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0056
Location:	LONG BRANCH POLICE DEPT	
Calibration File No.:	00665	Calib. Date: 04/04/2017
Certification File No.:	00666	Cert. No.: 00020
Linearity File No.:	00667	Lin. Date: 04/04/2017
Solution File No.:	00664	Lin. No.: 00020
Sequential File No.:	00667	Soln. Date: 04/03/2017
		Soln. No.: 00136
		File Date: 04/04/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.040%	Serial No.: DDRK S3-0005
Solution Control Lot:	16230	Expires: 09/19/2018
		Bottle No.: 0701
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.080%	Serial No.: DDRF S3-0009
Solution Control Lot:	16250	Expires: 09/27/2018
		Bottle No.: 1408
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.160%	Serial No.: DDCN-0054
Solution Control Lot:	16260	Expires: 10/03/2018
		Bottle No.: 0397

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:54D		
Control 1 EC	0.041%	12:55D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.040%	12:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:56D		
Control 2 EC	0.041%	12:57D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.040%	12:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:59D		
Control 3 EC	0.082%	12:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.079%	12:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:01D		
Control 4 EC	0.081%	13:02D	33.9°C	*** TEST PASSED ***
Control 4 IR	0.079%	13:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:03D		
Control 5 EC	0.162%	13:04D	34.0°C	*** TEST PASSED ***
Control 5 IR	0.160%	13:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:06D		
Control 6 EC	0.162%	13:07D	34.0°C	*** TEST PASSED ***
Control 6 IR	0.159%	13:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:08D		


All tests within acceptable tolerance.

Coordinator

Last Name: KLIMIK

First Name: DAVID

MI: W

Signature:  #7040

Badge No.: 7040

Date: 04/04/2017