

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00499 **Calib. Date:** 04/08/2015 **Calib. No.:** 00020
Certification File No.: 00500 **Cert. Date:** 04/08/2015 **Cert. No.:** 00015
Linearity File No.: 00501 **Lin. Date:** 04/08/2015 **Lin. No.:** 00015
Solution File No.: 00528 **Soln. Date:** 08/15/2015 **Soln. No.:** 00106
Sequential File No.: 00528 **File Date:** 08/15/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXN S3-0216
Control Solution %: 0.100% **Expires:** 01/16/2016
Solution Control Lot: 14A126 **Bottle No.:** 1400

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:37D		
Control 1 EC	-.---%	03:38D	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	03:38D	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JOYCE

First Name: PATRICK

MI: D

Signature: _____

Badge No.: 326

Date: 08/15/2015

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00499 Calib. Date: 04/08/2015 Calib. No.: 00020
Certification File No.: 00500 Cert. Date: 04/08/2015 Cert. No.: 00015
Linearity File No.: 00501 Lin. Date: 04/08/2015 Lin. No.: 00015
Solution File No.: 00522 Soln. Date: 07/17/2015 Soln. No.: 00105
Sequential File No.: 00522 File Date: 07/17/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0216
Control Solution %: 0.100% Expires: 01/06/2016
Solution Control Lot: 14A126 Bottle No.: 0284

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:59D		
Control 1 EC	0.096%	21:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:00D		
Control 2 EC	0.096%	21:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:02D		
Control 3 EC	0.096%	21:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OLSKI

First Name: GEORGE

MI: T

Signature: _____

Badge No.: 336

Date: 07/17/2015

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00499 **Calib. Date:** 04/08/2015 **Calib. No.:** 00020
Certification File No.: 00500 **Cert. Date:** 04/08/2015 **Cert. No.:** 00015
Linearity File No.: 00501 **Lin. Date:** 04/08/2015 **Lin. No.:** 00015
Solution File No.: 00517 **Soln. Date:** 06/08/2015 **Soln. No.:** 00104
Sequential File No.: 00517 **File Date:** 06/08/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXN S3-0216
Control Solution %: 0.100% **Expires:** 01/06/2016
Solution Control Lot: 14A126 **Bottle No.:** 0280

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:31D		
Control 1 EC	0.101%	23:32D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:33D		
Control 2 EC	0.100%	23:33D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:34D		
Control 3 EC	0.099%	23:35D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:36D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: UMANA

First Name: HECTOR

MI: D

Signature: 

Badge No.: 328

Date: 06/08/2015

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00473	Calib. Date:	10/24/2014
Certification File No.:	00474	Cert. Date:	10/24/2014
Linearity File No.:	00475	Lin. Date:	10/24/2014
Solution File No.:	00493	Soln. Date:	03/02/2015
Sequential File No.:	00493	File Date:	03/02/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0216
Solution Control Lot:	14A126	Expires:	01/06/2016
		Bottle No.:	0281

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:08S		
Control 1 EC	0.099%	01:08S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:09S		
Control 2 EC	0.099%	01:10S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:11S		
Control 3 EC	0.098%	01:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WALLS

First Name: CHRIS

MI:

Signature: _____

Badge No.: 314

Date: 03/02/2015