

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00593	Calib. Date:	04/22/2016
Certification File No.:	00594	Cert. Date:	04/22/2016
Linearity File No.:	00595	Lin. Date:	04/22/2016
Solution File No.:	00620	Soln. Date:	09/18/2016
Sequential File No.:	00620	File Date:	09/18/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0216
Solution Control Lot:	15H140	Expires:	08/04/2017
		Bottle No.:	1348

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:52D		
Control 1 EC	0.098%	19:53D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	19:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:53D		
Control 2 EC	0.098%	19:54D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:55D		
Control 3 EC	0.098%	19:55D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	19:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:56D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WAGNER

First Name: JOSHUA

MI: J

Signature: _____

PO J Wagner # 323

Badge No.: 323

Date: 09/18/2016