

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00779	Calib. Date:	05/30/2019
Certification File No.:	00780	Cert. Date:	05/30/2019
Linearity File No.:	00781	Lin. Date:	05/30/2019
Solution File No.:	00788	Soln. Date:	08/18/2019
Sequential File No.:	00788	File Date:	08/18/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0217
Solution Control Lot:	19060	Expires:	02/11/2021
		Bottle No.:	0688

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:19D		
Control 1 EC	0.101%	07:19D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:20D		
Control 2 EC	0.097%	07:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	07:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:22D		
Control 3 EC	0.098%	07:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	07:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CHAPARRO

First Name: SERGIO

MI: A

Signature: 

Badge No.: 352

Date: 08/18/2019