

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00779	Calib. Date:	05/30/2019
Certification File No.:	00780	Cert. Date:	05/30/2019
Linearity File No.:	00781	Lin. Date:	05/30/2019
Solution File No.:	00798	Soln. Date:	09/08/2019
Sequential File No.:	00798	File Date:	09/08/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0217
Solution Control Lot:	19060	Expires:	02/11/2021
		Bottle No.:	0627

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:28D		
Control 1 EC	0.098%	18:29D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:30D		
Control 2 EC	0.097%	18:31D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	18:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:32D		
Control 3 EC	0.097%	18:32D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORRELL

First Name: EVAN

MI: J

Signature: _____

Badge No.: 372

Date: 09/08/2019