

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0056  
**Location:** LONG BRANCH POLICE DEPT  
**Calibration File No.:** 00854 **Calib. Date:** 07/23/2020 **Calib. No.:** 00033  
**Certification File No.:** 00855 **Cert. Date:** 07/23/2020 **Cert. No.:** 00027  
**Linearity File No.:** 00856 **Lin. Date:** 07/23/2020 **Lin. No.:** 00027  
**Solution File No.:** 00866 **Soln. Date:** 10/08/2020 **Soln. No.:** 00182  
**Sequential File No.:** 00866 **File Date:** 10/08/2020

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDXN S3-0217  
**Control Solution %:** 0.100% **Expires:** 09/02/2022  
**Solution Control Lot:** 20390 **Bottle No.:** 0057

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 15:14D        |                               |                           |
| Control 1 EC      | 0.097%         | 15:15D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.097%         | 15:15D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 15:16D        |                               |                           |
| Control 2 EC      | 0.097%         | 15:16D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.098%         | 15:16D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 15:17D        |                               |                           |
| Control 3 EC      | 0.097%         | 15:18D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.098%         | 15:18D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 15:19D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: CARROLL

First Name: SHANE

MI: M

Signature: 

Badge No.: 363

Date: 10/08/2020