

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00830 **Calib. Date:** 02/18/2020 **Calib. No.:** 00032
Certification File No.: 00831 **Cert. Date:** 02/18/2020 **Cert. No.:** 00026
Linearity File No.: 00832 **Lin. Date:** 02/18/2020 **Lin. No.:** 00026
Solution File No.: 00839 **Soln. Date:** 04/08/2020 **Soln. No.:** 00173
Sequential File No.: 00839 **File Date:** 04/08/2020

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXN S3-0216
Control Solution %: 0.100% **Expires:** 02/11/2021
Solution Control Lot: 19060 **Bottle No.:** 0683

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:51D		
Control 1 EC	0.100%	06:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:53D		
Control 2 EC	0.098%	06:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:54D		
Control 3 EC	0.099%	06:55D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:56D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARROLL

First Name: SHANE

MI: M

Signature: Shane Carroll #363

Badge No.: 363

Date: 04/08/2020