

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00830 **Calib. Date:** 02/18/2020 **Calib. No.:** 00032
Certification File No.: 00831 **Cert. Date:** 02/18/2020 **Cert. No.:** 00026
Linearity File No.: 00832 **Lin. Date:** 02/18/2020 **Lin. No.:** 00026
Solution File No.: 00843 **Soln. Date:** 06/10/2020 **Soln. No.:** 00175
Sequential File No.: 00843 **File Date:** 06/10/2020

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXN S3-0216
Control Solution %: 0.100% **Expires:** 02/11/2021
Solution Control Lot: 19060 **Bottle No.:** 0633

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:15D		
Control 1 EC	0.101%	04:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:16D		
Control 2 EC	0.100%	04:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:18D		
Control 3 EC	0.098%	04:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	04:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SAMOL

First Name: GEORGE

MI: L

Signature:

 #330

Badge No.: 330

Date: 06/10/2020