

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0056  
**Location:** LONG BRANCH POLICE DEPT  
**Calibration File No.:** 00878 **Calib. Date:** 01/12/2021 **Calib. No.:** 00034  
**Certification File No.:** 00879 **Cert. Date:** 01/12/2021 **Cert. No.:** 00028  
**Linearity File No.:** 00880 **Lin. Date:** 01/12/2021 **Lin. No.:** 00028  
**Solution File No.:** 00890 **Soln. Date:** 04/02/2021 **Soln. No.:** 00191  
**Sequential File No.:** 00890 **File Date:** 04/02/2021

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDXN S3-0216  
**Control Solution %:** 0.100% **Expires:** 09/02/2022  
**Solution Control Lot:** 20390 **Bottle No.:** 0119

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 04:36S        |                               |                           |
| Control 1 EC      | 0.099%         | 04:36S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.099%         | 04:36S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 04:37S        |                               |                           |
| Control 2 EC      | 0.098%         | 04:38S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.099%         | 04:38S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 04:39S        |                               |                           |
| Control 3 EC      | 0.098%         | 04:40S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.098%         | 04:40S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 04:40S        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: CARROLL

First Name: SHANE

MI: M

Signature:

*PO Shane Carroll #363*

Badge No.: 363

Date: 04/02/2021

# Calibrating Unit

## New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C  
Location: LONG BRANCH POLICE DEPT.  
Calibration File No.: 00878  
Certification File No.: 00879  
Linearity File No.: 00880  
Solution File No.: 00885  
Sequential File No.: 00885  
Calib. Date: 01/12/2021  
Cert. Date: 01/12/2021  
Lin. Date: 01/12/2021  
Soln. Date: 02/24/2021  
File Date: 02/24/2021  
Serial No.: ARXA-0056  
Calib. No.: 00034  
Cert. No.: 00028  
Lin. No.: 00028  
Soln. No.: 00189

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 20390  
Model No.: CU-34  
Serial No.: DDXN S3-0216  
Expires: 09/02/2022  
Bottle No.: 0058

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 02:35S        |                               |                           |
| Control 1 EC      | 0.100%         | 02:36S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.098%         | 02:36S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 02:37S        |                               |                           |
| Control 2 EC      | 0.098%         | 02:37S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.098%         | 02:37S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 02:38S        |                               |                           |
| Control 3 EC      | 0.098%         | 02:39S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.099%         | 02:39S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 02:40S        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SAMOL

First Name: GEORGE

MI: L

Signature:

Badge No.: 330

Date: 02/24/2021

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0056  
Location: LONG BRANCH POLICE DEPT  
Calibration File No.: 00878 Calib. Date: 01/12/2021 Calib. No.: 00034  
Certification File No.: 00879 Cert. Date: 01/12/2021 Cert. No.: 00028  
Linearity File No.: 00880 Lin. Date: 01/12/2021 Lin. No.: 00028  
Solution File No.: 00881 Soln. Date: 01/12/2021 Soln. No.: 00188  
Sequential File No.: 00881 File Date: 01/12/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0216  
Control Solution %: 0.100% Expires: 09/02/2022  
Solution Control Lot: 20390 Bottle No.: 0123

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 11:16S        |                               |                           |
| Control 1 EC      | 0.100%         | 11:17S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.100%         | 11:17S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 11:18S        |                               |                           |
| Control 2 EC      | 0.098%         | 11:19S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.099%         | 11:19S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 11:20S        |                               |                           |
| Control 3 EC      | 0.098%         | 11:21S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.099%         | 11:21S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 11:22S        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXAP2-114 DL

### Changed By:

Last Name: LUTZ

First Name: DENNIS

MI: J

Signature: Tpr I Lutz 7045

Badge No.: 7045

Date: 01/12/2021