

Calibrating Unit New Standard Solution Report

22LB 12752

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00939 **Calib. Date:** 12/27/2021 **Calib. No.:** 00036
Certification File No.: 00940 **Cert. Date:** 12/27/2021 **Cert. No.:** 00030
Linearity File No.: 00941 **Lin. Date:** 12/27/2021 **Lin. No.:** 00030
Solution File No.: 00961 **Soln. Date:** 04/25/2022 **Soln. No.:** 00211
Sequential File No.: 00961 **File Date:** 04/25/2022

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXN S3-0216
Control Solution %: 0.100% **Expires:** 09/01/2023
Solution Control Lot: 21360 **Bottle No.:** 0293

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:02D		
Control 1 EC	0.097%	11:03D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:04D		
Control 2 EC	0.096%	11:05D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	11:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:06D		
Control 3 EC	0.096%	11:06D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: VECCHIONE

First Name: GARY

MI: M

Signature: _____

Badge No.: 343

Date: 04/25/2022